

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

> Mechanical and Electrical Testing (As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen

President/Operations Manager

Perry Johnson Laboratory Accreditation, Inc. (PJLA) 755 W. Big Beaver, Suite 1325 Troy, Michigan 48084

Initial Accreditation Date:

Issue Date:

Expiration Date:

November 2, 2015

November 2, 2015

February 28, 2018

Accreditation No.:

Certificate No.:

88632

L15-364

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlabs.com



Certificate of Accreditation: Supplement

MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

FIELD	ITEMS, MATERIALS	SPECIFIC TESTS OR	SPECIFICATION,	RANGE (WHERE
OF TEST	OR PRODUCTS TESTED	PROPERTIES MEASURED	STANDARD METHOD OR TECHNIQUE USED	APPROPRIATE) AND DETECTION LIMIT
Electrical F	Medical Devices	Electrochemical	ASTM F2129	N/A
		Corrosion	ASTN F3044	
			ASTM G59	
			ASTM G71	
			Lab Developed Method	
			ECOR-001	
		Electrosurgical	ANSI/AAMI/IEC 60601-	
		Accessory Electrical	2-2 Section 201.8.8.3.101	
		Testing	ANSI/AAMI/IEC 60601-	
			2-2 Section 201.8.8.3.102	
			ANSI/AAMI/IEC 60601-	
			2-2 Section 201.8.8.3.103	
			ANSI/AAMI/IEC 60601-	
			2-2 Section 201.8.8.3.104	
			IEC 60601-2-18 Section	
			201.11.101.2	
			Lab Developed Method	
			ECT-491	
Mechanical F		*MRI Testing	ASTM F2052, F2119,	
		(Displacement Force,	F2182, F2213;	
		Torque, RF Heating and	Lab Developed Method	
		Image Artifact)	MRI-400	
		Immersion Corrosion	ASTM F1980	
			ASTM G 31	
			ASTM 1089	
			Lab Developed Method	
			ACOR-716	
		**Pulsatile Fatigue	ASTM F2477	
			ASTM F3036	
			ASTM E739	
			Lab Developed Method	
			FATG -320	
		Tensile (up to 10000N)	BS EN 1617	10 KN
		Stress/Strain	BS EN 1618 Annex B	
		Maximum Force	ASTM E8/E8M	
			JIS T 3213	
			ASTM D412	
			BS EN 1615 Annex F	
			JIS T 3247	
			ISO 7864	
			Lab Developed Methods	
			PULT-204, -210,	
		Torque	ASTM A938;	±2.8 N⋅m
			Lab Developed Method	
			TORQ-553	



Certificate of Accreditation: Supplement

MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical F	Medical Devices	Particulate Matter Generation	ASTM F2734 ASTM F2942 Lab Developed Methods PART -02 ^{††} , -03, -04, -05	2 μm to 400 μm
		**Simulated Use	ASTM F2081 ASTM F2079 ANSI/AAMI VP20 - 8.1 Lab Developed Method SIM-01	N/A
		**†Physical Bench Examination and Mechanical Bench Measurements	ISO 5084 ASTM D1777 Lab Developed Methods MED-007, MEAS-825, -829, -830, -832	
		Fatigue (Flat Plate, Bending, Axial, Torsional and Sling Radial)	ASTM F2942, ASTM E739 Lab Developed Methods FATG-401, -500, -700, -800, -900	± 12.7 mm ± 180° ± 2 225 N
		Flow	ASTM F1828 ANSI/AAMI VP20 Lab Developed Methods FLOW-100, -302	N/A
		Pressure	ANSI/AAMI VP20 Section 8.3.3.3 ISO 13938-1 Lab Developed Method PRES-01	-103 kPa to +207 KPa
		Securement	ASTM F2394 Lab Developed Method MEAS-100	N/A
		Radial Force	ASTM 3067 (segmented head apparatus method) Lab Developed Method RF-300	0.5 mm to 55 mm 0 N to 225 N
		**Radiopacity	ASTM F640 Lab Developed Method RAD-01	N/A
		Compression	ASTM E9 ASTM D695 ASTM F2606 Lab Developed Method COMP-220	10 KN



Certificate of Accreditation: Supplement

MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical F	Medical Devices	Abrasion (Martindale) ^{††}	ASTM D4966 ISO 12947-1 ISO 12947-3 Lab Developed Method MART -100	N/A
		Yarn Testing	ASTM D2259 ASTM D1907 ASTM D2256 ASTM D2259 ASTM D1779 ASTM D1423 Lab Developed Method Yarn-100	
	Tubular vascular prostheses Intravascular catheters- Sterile and single-use catheters	See pg. 4 for details See pg. 5 for details	ISO 7198 ISO 10555-1	
	Angiographic catheters	See pg. 5 for details	ISO 10555-2	
	Balloon dilatation catheters	See pg. 5 for details	ISO 10555-4	
	Endovascular prostheses	See pg. 6 for details	ISO 25539-1	
	Vascular stents	See pg. 7 for details	ISO 25539-2	
	Vena cava filters	See pg. 8 for details	ISO 25539-3	
	Sterile single-use intravascular introducers, dilators and guidewires	See pg. 9 for details	ISO 11070	
	Cardiovascular implants -	Visual Inspection	ISO 7198 section 8.1	
	Tubular vascular prostheses	Liquid Leakage Porosity Water Permeability Integral Water Permeability/Leakage Water Entry Pressure	ISO 7198 section 8.2	
		Strength Testing Circumferential Tensile Strength Longitudinal Tensile Strength Burst Strength Strength after Repeated Puncture	ISO 7198 section 8.3	





Certificate of Accreditation: Supplement

MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical F	Cardiovascular implant Tubular	Usable Length	ISO 7198 section 8.4	N/A
		Relaxed Internal Diameter	ISO 7198 section 8.5	
	vascular prostheses	Pressurized Internal Diameter	ISO 7198 section 8.6	
		Wall Thickness	ISO 7198 section 8.7.4.2	
		Suture Retention Strength	ISO 7198 section 8.8	
		Kink Diameter/Radius	ISO 7198 section 8.9	
		Dynamic Compliance	ISO 7198 section 8.10	
		Visual Inspection	ISO 7198 section 8.1	
	Intravascular catheters- Sterile and single-use catheters Part 1: General requirements Part 2: Angiographic catheters Part 4: Balloon dilatation catheters	Test method for corrosion resistance	ISO10555-1 Annex A	
		Method for determining peak tensile force	ISO10555-1 Annex B	
		Test Method for liquid leakage under pressure	ISO 10555-1 Annex C	
		Test method for air leakage into hub assembly during aspiration	ISO 10555-1 Annex D	
		Determination of flow rate	ISO 10555-1 Annex E	
		Test for burst pressure under static conditions	ISO 10555-1 Annex F	
		Power injection test for flow rate and device pressure	ISO 10555-1 Annex F	
		Test for freedom from leakage and damage under high static pressure conditions	ISO 10555-2 Annex A	
		Test for balloon rated burst pressure	ISO 10555-4 Annex A	
		Balloon Fatigue test for freedom from leakage and damage on inflation	ISO 10555-4 Annex B	
		Test for balloon deflation time	ISO 10555-4 Annex C	
		Test for balloon diameter to inflation pressure	ISO 10555-4 Annex D	





MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical F	Cardiovascular	**Dimension verification and	ISO 25539-1/Amd.1	N/A
	implants- Endovascular devices Part 1:Endovascular	component dimensional		
		compatibility Profile/diameter test	ISO 25539-1/Amd.1	
	prostheses			
	prostneses	Assessment of hemostasis	ISO 25539-1/Amd.1	
		**Simulated use models	ISO 25539-1/Amd.1	
		**Visibility	ISO 25539-1/Amd.1	
		Force to deploy	ISO 25539-1/Amd.1	
		Balloon inflation and deflation time	ISO 25539-1/Amd.1	
		Balloon rated burst pressure	ISO 25539-1/Amd.1	
		Balloon volume to burst	ISO 25539-1/Amd.1	
		Balloon rated fatigue	ISO 25539-1/Amd.1	
		Bond Strength	ISO 25539-1/Amd.1	
		Torsional bond strength	ISO 25539-1/Amd.1	
		Tubing longitudinal tensile strength	ISO 25539-1/Amd.1	
		Dimensional verification	ISO 25539-1/Amd.1	
		Implant diameter to balloon inflation pressure	ISO 25539-1/Amd.1	
		Implant length to diameter relationship	ISO 25539-1/Amd.1	
		Recoil	ISO 25539-1/Amd.1	
		Integral water permeability/leakage	ISO 25539-1/Amd.1	
		Water entry pressure	ISO 25539-1/Amd.1	
		Water permeability	ISO 25539-1/Amd.1	
		Burst/circumferential strength	ISO 25539-1/Amd.1	
		Crush Resistance	ISO 25539-1/Amd.1	1
		Flex/Kink	ISO 25539-1/Amd.1	1
		Local compression	ISO 25539-1/Amd.1	
		Longitudinal tensile strength	ISO 25539-1/Amd.1	1





Certificate of Accreditation: Supplement

MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical F	Cardiovascular	Migration resistance	ISO 25539-1/Amd.1	N/A
	implants- Endovascular devices Part 1:Endovascular	Pull test for modular components	ISO 25539-1/Amd.1	
		Radial force	ISO 25539-1/Amd.1	
	prostheses	Strength after repeated puncture	ISO 25539-1/Amd.1	
		Strength of graft to stent/attachment system bond	ISO 25539-1/Amd.1	
		Corrosion assessment	ISO 25539-1/Amd.1 Annex D5.3.18	
		Fatigue and durability test (pulsatile)	ISO 25539-1/Amd.1	
	Cardiovascular implants- Endovascular devices Part 2:Endovascular prostheses	**Dimension verification and component dimensional compatibility	ISO 25539-2 Annex	
		Profile/diameter test	ISO 25539-2 Annex]
		Assessment of hemostasis	ISO 25539-2 Annex	
		**Simulative use	ISO 25539-2 Annex	
		**Visibility	ISO 25539-2 Annex	
		Force to deploy	ISO 25539-2 Annex	
		Balloon inflation and deflation time	ISO 25539-2 Annex	
		Balloon rated burst pressure	ISO 25539-2 Annex	
		Balloon rated fatigue	ISO 25539-2 Annex	
		Bond Strength	ISO 25539-2 Annex	
		Torsional bond strength	ISO 25539-2 Annex	
		Stent diameter to balloon inflation pressure	ISO 25539-2 Annex	
		Dimensional verification and stent length to diameter relationship	ISO 25539-2 Annex	
		Recoil	ISO 25539-2 Annex	





MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical F	Cardiovascular implants- Endovascular devices –	Crush resistance with radially applied load	ISO 25539-2 Annex	N/A
	Part 2:Endovascular prostheses	Crush resistance with parallel plates	ISO 25539-2 Annex	
	r	Flex/kink	ISO 25539-2 Annex	-
		Local compression	ISO 25539-2 Annex	-
		Radial Force	ISO 25539-2 Annex	-
		Corrosion assessment	ISO 25539-2 Annex	-
		Fatigue durability test	ISO 25539-2 Annex	-
		Dislodgment force	ISO 25539-2 Annex	-
		Dogboning	ISO 25539-2 Annex	-
		Profile effect/flaring	ISO 25539-2 Annex	-
		Stent free surface area and	ISO 25539-2 Annex	-
		stent outer surface area		-
		Acute coating integrity	ISO 25539-2 Annex	
	Cardiovascular implants- Endovascular devices – Part 3:Vena cava filters	**Dimension verification and component dimensional compatibility	ISO 25539-3 Annex D5.1.1 and D5.5.1	
		**Simulative use	ISO 25539-3 Annex D.5.1.2	
		Force to deploy	ISO 25539-3 Annex D.5.1.3	
		**Visibility	ISO 25539-3 Annex D.5.1.4 and D5.5.4	
		Clot trapping	ISO 25539-3 Annex D.5.2.1	
		Fatigue/durability	ISO 25539-3 Annex D.5.2.2	
		Filter dimensional verification	ISO 25539-3 Annex D.5.2.3	
		Filter tensile strength	ISO 25539-3 Annex D.5.2.4	
		Migration resistance	ISO 25539-3 Annex D.5.2.5	
		Radial force	ISO 25539-3 Annex D.5.2.6	
		Visual inspection	ISO 25539-3 Annex D.5.2.8	





MED Institute, Inc. 1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical F	Cardiovascular implants-	Tensile strength	ISO 25539-3 Annex	N/A
	Endovascular devices - Part		D.5.3.1, D.5.4.3, D.5.6.1	
	3:Vena cava filters		and D.5.7.3	
		Torsional bond strength	ISO 25539-3 Annex	
			D.5.3.2, D.5.4.4,	
			D.5.6.2 and D.5.7.4	
		Catheter burst	ISO 25539-3 Annex	
			D.5.4.1 and D.5.7.1	
		Power injection	ISO 25539-3 Annex	
			D.5.4.2 and D.5.7.2	
		Simulated use (endovascular	ISO 25539-3 Annex	
		retrieval/conversion system)	D.5.5.2	
		Force to retrieve/convert	ISO 25539-3 Annex	
			D.5.5.3	
	Sterile single-use intravascular	Test method for corrosion	ISO 11070 Annex B	
	introducers, dilators and	resistance		
	guidewires	Determination of force at	ISO 11070 Annex C	
		break of introducer		
		catheters, sheath introducers		
		and dilators		
		Test for liquid leakage from	ISO 11070 Annex D	
	/	sheath introducers under		
		pressure	75.2 11.8 E.	
		Test for liquid leakage	ISO 11070 Annex E	
		through haemostasis valves		
		of sheath introducers	700 11070 1	
		Test for fracture of guide	ISO 11070 Annex F	
		wires	ICO 11070 A II	
		Test for resistance of guide	ISO 11070 Annex H	
		wires damage by flexing	100 11000 A	
		Determination of strength of	ISO 11070 Annex I	
		union of needle hub and		
		needle		





MED Institute, Inc.

1330 Win Hentschel Blvd. STE 100, West Lafayette, IN 47906 Contact: Justin Metcalf Phone: 765-463-1633

Accreditation is granted to the facility to perform the following testing:

- 1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer F would mean that the laboratory performs this testing at its fixed location.
- 2. *MRI testing is conducted using clinical scanners at offsite facilities.
- **Radiographic activity, when performed as part of these tests, is performed using equipment at:
 Purdue University
 Lynn Hall
 625 Harrison St.
 West Lafayette, IN 47906
- 4. † SEM and High resolution X-ray, when performed, is performed using equipment at:
 CRI
 1 Geddes way
 West Lafayette, IN 47906

5. †† Testing is performed using equipment at: CRI1 Geddes wayWest Lafayette, IN 47906

